

UL Standard Declaration of Conformity

Company: Chemyx Inc.
10905 Cash Road,
Stafford, TX 77477 USA

Product Name: Fusion 100 XT Metering Pump

Model Name: F100XT

This is a certificate of conformity to certify that the bearer has successfully completed the requirements of the above scheme which include the testing of products, the initial assessment, and are subject to continuing annual assessments of their compliance and testing of samples of products taken from production (as applicable to the scheme) and has been registered within the scheme for the products detailed with the applicable UL harmonized standards; CSA harmonized standards and other normative requirements:

Testing Standards: IEC 61010-1:2012 Ed.3+R:06 Jun2023, Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

EN IEC 61326-1:2021 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use Part 1:General Requirements (R2022) [CSA C22.2#61010-1-12:2012 Ed.3+U1;U2;A1]

Conforms To: ANSI/UL STD. 61010-1

Certified To: CAN/CSA STD. C22.2 No 61010-1

Testing Performed By: Intertek Testing Services NA Inc.
545 East Algonquin Road, Arlington Heights, IL 60005



Signature on Behalf of Manufacture
or Authorized Representative

A rectangular box containing a handwritten signature in black ink, which appears to be 'Jeff Wu'.

Name:
Title:
Date:

Jeff Wu
Executive Director
3/5/2025